RADIOACTIVE SOURCES

Model RN-1025, TH-1025 and RNC sources are flow-through sources that provide a convenient means of generating either radon or thoron gas

Sources:

Calibrated gas sources were developed by Pylon to satisfy the need for a convenient and accurate means of checking field instrumentation. These sources feature reliable rate of emanation, high accuracy and ease of use. They are widely used for field checking of radon and thoron measuring instruments.

Flow-Through Sources:

The Pylon Models RNC, RN-1025 and TH-1025 are convenient flow-through radioactive gas sources, ideally suited for continuous or intermittent applications in environmental chambers, radon and thoron enclosures and for the checking of instruments. They contain a dry radium (Ra-226) or thorium (Th-228) source which provides calibrated quantities of radon (Rn-222) or thoron (Rn-220) for laboratory or field use.

Applications:

Radon or thoron monitor/detector checks

Features:

- ANSI N538-1979 safety standard compliant
- High accuracy
- Easy to use
- Continuous operation
- Constant rate of emanation

Theory of Operation:

The classical method of producing radon gas involves degassing a solution containing radium salts, a process requiring careful attention to detail if accurate results are to be obtained. The equipment is fragile and generally not suitable for field use. There is also the attendant risk of spills and contamination.



RN-1025/TH-1025



RNC

Pylon sources provide a convenient means of generating calibrated quantities of either radon or thoron gas. The source material is in a dry powder form which reliably emanates 100% of the gas produced.



RADIOACTIVE SOURCES

Specifications:

| <u>opcomoationo.</u> | | | | |
|--|-------------------------------|---------------|--------------------|-----------|
| | RN-1025 | TH-1025 RNC | | |
| Parent nuclide: | Ra-226 | Th-228 | Ra-226 | |
| Nominal activity: | See PN Table | See PN Table | 0.002 (0.074) | uCi (kBq) |
| Activity Tolerance: | -10 / +25-10 / +25-10 / +25% | | | |
| Calibration Accuracy ¹ : | ± 4 | ± 8 | ± 4 | % |
| Housing Material: | Aluminum | Aluminum | Aluminum | |
| Maximum Flow Rate | 10 | 10 | 10 | lpm |
| Operating Temperature Range: | 0 to +50 0 to +50 0 to +50 °C | | | |
| | (+32 to +122) | (+32 to +122) | (+32 to +122) | (°F) |
| Storage Temperature Range: | -20 to +75 | -20 to +75 | -20 to +75 | °C |
| _ | (-4 to +167) | (-4 to +167) | (-4 to +167) | (°F) |
| Relative Humidity Range ² : | 0 to 90 | 0 to 90 | 0 to 90 | % |
| Length 3: | 15.2 (6) | 15.2 (6) | - | cm (in.) |
| Width ³ : | 48.3 (19) | 48.3 (19) | - | cm (in.) |
| Height ⁴ _: | 10.2 (4) | 10.2 (4) | - | cm (in.) |
| Height ⁵ .: | 13.3 (5.3) | 13.3 (5.3) | - | cm (in.) |
| Weight 4: | 1.6 (3.5) 1.6 (3.5) |) - | kg (lb.) | |
| Weight ⁵ : | 13.6 (30) | 13.6 (30) | - | kg (lb.) |
| Diameter ⁶ : | - | - | 6.7 (2.65) | cm (in.) |
| Height ⁶ : | - | - | 9.1 (3.6) cm (in.) | |
| Weight ⁶ : | - | - | 525 (1.16) | g (lb.) |
| 1 At a 4 Confidence Level | | | | |

 $^{^{1}}$ At a 1 $_{\odot}$ Confidence Level.

Ordering Information:

| Model | Nominal Source Activity | Order Part Number |
|-------------------------------|---|----------------------|
| RN-1025-Custom-S 7 | < 2000 kBq (54 μCi) | 6202074 |
| RN-1025-Custom-L ⁷ | ≥ 2000 kBq (54 uCi) & < 4070 kBq (110 uCi) | 6202075 |
| TH-1025-Custom-S 7 | < 2000 kBq (54 µCi) | 6202094 |
| TH-1025-Custom-L ⁷ | ≥ 2000 kBq (54 uCi) & < 4070 kBq (110 uCi) | 6202095 |
| RNC | 0.002 μCi (74 Bq) | A206324 |

⁷ Activity is specified by the Customer.



Source Panel made with several RN-1025 sources in parallel to provide an array of radon concentration for your chamber.

THESE SOURCES CONTAIN RADIOACTIVE MATERIALS, PYLON REQUIRES A COPY OF THE USER'S RADIOISOTOPE LICENCE PRIOR TO SHIPPING THESE PRODUCTS.

Specifications subject to change without notice.

Trademarks are the properties of their respective holders. All Rights Reserved. Datasheet: 123 Rev 4

² Non-Condensing.

³ 1025.

⁴ Small Body 1025 (< 54 uCi (2000 kBq))

⁵ Large Body 1025 (≥ 54 uCi (2000 kBq))

[·] Values are nominal.